

# Liofilchem®

## Certificate of Analysis

Page 1 of 1

| Product                                         | Batch     | Production date | Expiration date |
|-------------------------------------------------|-----------|-----------------|-----------------|
| <b>Voriconazole VO</b><br><b>0.002-32 µg/mL</b> | 011625093 | 16.01.2025      | 2028.01.16      |

Ref.

92150 – 921500 – 921501

### Antimicrobial Susceptibility Testing

Tested according to in-house methodology for antifungal agents

| Control strains                            | Medium        | Inoculum                   | Incubation                   | Expected Results<br>MIC range (mg/L) | Results<br>MIC(mg/L) |
|--------------------------------------------|---------------|----------------------------|------------------------------|--------------------------------------|----------------------|
| <i>Candida krusei</i><br>ATCC® 6258        | R.P.M.I. Agar | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>24-48 h | 0.25-1                               | 0.25                 |
| <i>Candida parapsilosis</i><br>ATCC® 22019 | R.P.M.I. Agar | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>24-48 h | 0.016-0.064                          | 0.047                |
| <i>Candida albicans</i><br>ATCC® 90028     | R.P.M.I. Agar | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>24-48 h | 0.004-0.016                          | 0.016                |

### Batch Release

Approved

Date

24.01.2025

Signature

Quality Control  
(This document has been  
established electronically and  
is valid without signature)

The results reported were obtained at the time of release.