

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
Daptomycin DAP 0.016-256 µg/mL	110424083	04.11.2024	2027.11.04

Ref.

92145 – 921450 – 921451

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.12–1	0.19
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–4	1
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar (Sheep blood 5%)	0.5 McFarland in broth	35 ± 2°C, 5% CO ₂ 20-24 h	0.06–0.5	0.12

Batch Release

Approved

Date

14.11.2024

Signature

Quality Control
(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.