



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® C. difficile BT

Catalogue Part No.: GXCDIFFBT-CE-10

Kit Lot No.: 1001477726

Cartridge Lot No.: 37901

Kit Expiration Date: 2026-09-06

Legal Manufacturer

Cepheid AB
Röntgenvägen 5
SE-171 54 Solna,
Sweden

Manufacturing Facility

Cepheid AB
Röntgenvägen 5
SE-171 54 Solna,
Sweden

Functional Testing

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
Positive Control	Toxigenic C.diff POS;Binary Toxin POS;027 PRESUMPTIVE POS	Passed
Negative Control	Toxigenic C.diff NEG;Binary Toxin NEG;027 NEG	Passed

☒ If checked this document is produced electronically and valid without a wet signature.

Alexander Avramidis
Alexander Avramidis (Mar 26, 2025 11:20 GMT+1)

Signature of Quality Assurance

Date

Name: Alexander Avramidis

Title: QA Analyst