

CERTIFICATE OF ANALYSIS SI 390.900 – URO-QUICK SCREENING KIT

Code

MIC_COA_SI390900#SCR250149

PRODUCT NAME	URO QUICK SCREENING KIT	
REF NUMBER	SI 390.900	
LOT	SCR250149	
MANUFACTURING DATE (yyyy-mm-dd)	2025-05-15	
EXPIRATION DATE (yyyy-mm-dd)	2026-07-14	

Tests	Specifications	Results				
Visual inspection						
Appearance of the liquid culture medium Appearance of the vial	Limpid, yellow, transparent liquid Green ring cap, pierceable septum, white stripe, presence of the stirrer, absence of defects	OK				
Microbiological activity						
Staphylococcus epidermidis NCTC 11047¹ Growth detection with inoculum ≤ 100 CFL		OK				
Microbiological state						
72 hours of incubation at 37°C Absence of turbidity of the liquid culture media ³		OK				

¹ Equivalent to CECT232; ATCC14990

The sampling method and quality controls have been performed in accordance with the procedures in force.

Analysis outcome:	☑ Compliant □ Not compliant			
Lot approved on	2025-05-19	By: (Quality Assurance)	Elens	low zzono

©Copyright [2024]. This document is exclusive property of ALIFAX S.r.I, confidential and not freely disclosable. All rights reserved.

ALIFAX S.r.l.

Registered and head office via Petrarca, 2/1 - 35020 Polverara PD Italy Tel. (+39) 049 0992000 Fax (+39) 049 5855434 info@alifax.com - info.alifax@pec.alifax.com R&D, Design and Manufacturing Dept via Merano, 30,33045 Nimis UD Italy Tel. (+39) 0432 197900 Fax (+39) 0432 547378 info.udine@alifax.com - info.udine@pec.alifax.com

²Inoculum concentration guaranteed using certified standards.

³Visual inspection of the liquid culture medium after incubation