

CERTIFICATE OF ANALYSIS

Product Name: Streptococcal Grouping Reagent Latex Suspension Group G	Lot Number: C1335B
Product Code: PL.036	Expiration Date: 2026-12-31

Additional Information

Storage Conditions: 2-8°C
 Instructions for Use: Refer to pro-lab.com/resources for current version

Quality Control Testing

Appearance	Reaction	Result
Homogenous suspension with no agglutination	N/A	Satisfactory
Organisms		
<i>Streptococcus</i> sp. ATCC 12394 (Group G)	Positive	PASS
<i>Streptococcus</i> Group G - Clinical Isolate	Positive	PASS
<i>Streptococcus pyogenes</i> ATCC 19615 (Group A)	Negative	PASS
<i>Streptococcus agalactiae</i> ATCC 12386 (Group B)	Negative	PASS
<i>Streptococcus equisimilis</i> ATCC 12388 (Group C)	Negative	PASS
<i>Enterococcus faecalis</i> ATCC 49477 (Group D)	*	PASS
<i>Streptococcus</i> sp. ATCC 12392 (Group F, type 2)	Negative	PASS
<i>Enterococcus faecalis</i> ATCC 19433	Negative	PASS
<i>Enterococcus faecalis</i> ATCC 49478	Negative	PASS
<i>Enterococcus faecalis</i> ATCC 29212	Negative	PASS
<i>Listeria monocytogenes</i> NCTC 11994	Negative	PASS
Control		
Polyvalent Positive Control (PL.040)	Positive	PASS

*Cross reaction may be present.

Release

2025-04-29	
Date	Quality Control

Pro-Lab Diagnostics U.K.

3 Bassendale Road, Bromborough, Wirral,
 Merseyside CH62 3QL
 United Kingdom
 Tel.: 0151 353 1613
 www.pro-lab.co.uk
 E-mail: uksupport@pro-lab.co.uk

Pro-Lab Diagnostics Canada

20 Mural Street, Unit #4
 Richmond Hill, Ontario, L4B 1K3
 Canada
 Tel.: 905-731-0300
 www.pro-lab.com
 E-mail: support@pro-lab.com

Pro-Lab Diagnostics U.S.A.

1301 Blue Ridge Drive, Suite #101
 Georgetown, Texas 78626-3269
 United States
 Tel.: 512-832-9145
 www.pro-lab-direct.com
 E-mail: support@pro-lab.us