

EU DECLARATION OF CONFORMITY

Code: MOL_DECO_SI17010103_MMGRAMPOSSTAPH-1-0_EN.docx

Date:

2022-05-18

The Manufacturer:

Manufacturer Name/Registered trade Name	ALIFAX S.R.L.
Address	Via Merano 30
	33045 NIMIS (UD)
	ITALY

declares on his own responsibility that the following devices:

Product Name and	MM GRAM POS STAPH	SI 1701.0103
Product Code		SI 1701.0103/L
Intended Purpose	professional clinical laboratory users, sequences specific to Gram-positive sequences associated with their non-starting from a positive blood culture of STAPH cartridge contains all the nethrough a multiplex Real Time PCR MOUSE SYSTEM. MM GRAM POS STAPH analyzes the aureus, S. epidermidis, S. haemolytic simulans, S. saprophyticus, S. xylosus, The negative result for the MecA, MecC not exclude the presence of other drug MM GRAM POS STAPH provides a reswith Gram-positive Staphylococcus material to methicillin and vancomycin. The MM GRAM POS STAPH test is integrated in the list of identifiable to included in the list of identifiable to independent tests are necessary.	tative in vitro diagnostic test, intended for for the multiplex identification of nucleic acid Staphylococcus bacteria and/or nucleic acid susceptibility to methicillin and vancomycin, of Gram-positive bacteria. The MM GRAM POS cessary reagents to perform one single test analysis in combination with MOLECULAR are following targets: Staphylococcus spp, S. cus, S. lugdunensis, S. sciuri, S. hominis, S. MecA, MecC, Scc mec-orfX, VanA and VanB. C., Scc mec-orfX, VanA and VanB targets does are resistance mechanisms. Sult to support the diagnosis of blood infection incroorganisms and/or their non-susceptibility and for combined use with other clinical and a evaluation defined and regulated by each STAPH does not replace traditional methods of biblity testing.
Risk Class	of the Directive 98/79/EC"	are neither in List A nor in List B of Annex II

complies with the Essential Requirements of the Directive 98/79/EC related to the In Vitro Diagnostic Medical Devices (IVDD).

We declare also that the device has been designed and manufactured in conformity to the below:

Harmonized Standards	EN ISO 13485:2016/A11:2021: Medical devices - Quality management systems - Requirements for
	regulatory purposes
	EN ISO 14971:2019: Medical devices - Application of risk management to medical devices
	EN ISO 15223-1:2021: Medical devices — Symbols to be used with medical device labels, labelling and
	information to be supplied
	EN ISO 18113-1:2011: In vitro diagnostic medical devices - Information supplied by the manufacturer –
	Part1 : Terms, definitions and general requirements
	EN ISO 18113-2:2011: In vitro diagnostic medical devices - Information supplied by the manufacturer –
	Part 2: In vitro diagnostic reagents for professional use
	EN 13612:2002/AC:2002: Performance evaluation of in vitro diagnostic medical devices
	EN ISO 23640:2015 In vitro diagnostic medical devices - Evaluation of stability of in vitro diagnostic
	reagents



EU DECLARATION OF CONFORMITY

Code: MOL_DECO_SI17010103_MMGRAMPOSSTAPH-1-0_EN.docx Date: 2022-05-18

Technical Standards	IEC 62366-1:2015+A1:2020: Application of usability engineering to medical devices IEC TR 62366-2:2016: Guidance on the application of usability engineering to medical device
Other	Directive 2015/863/EU (RoHS3)

The devices have been CE Marked as IVD Medical Device/Devices according to Annex III of the Directive 98/79/EC as they are neither in List A nor in List B of Annex II.

Any change made to the devices without the prior consent of the Manufacturer will automatically void the present Declaration of Conformity.

Place and date of issue: NIMIS (UD), 2022-05-18

Name and function: Camillo Galiano, Managing Director

Signature: