

Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

sifin diagnostics gmbh

Berliner Allee 317-321
13088 Berlin, Germany

T: +49 30 700 144-0
F: +49 30 700 144-30
info@sifin.de
www.sifin.de

Product	Anti-Shigella II Group A S. dysenteriae type 1-10
Catalogue no.	TS1821
Package size	1 ml
Lot no.	5860824
Expiry date	2029-07
Intended use	Anti-Shigella II records the serotypes of <i>Shigella dysenteriae</i> (type 1 to 10) in slide agglutination.
Specificity	No agglutination with the serotypes of <i>Shigella flexneri</i> , <i>Shigella boydii</i> (type 1 to 15) and <i>Shigella sonnei</i> (Phase I and Phase II).
Preservatives	Sodium azide 0.9 mg/ml
Storage	2 °C...8 °C

This document has been established electronically and is valid without signature.
The batch release was done in accordance with SOP QMV 7.