

# Liofilchem®

## Certificate of Analysis

Page 1 of 1

| Product                          | Batch     | Production date | Expiration date |
|----------------------------------|-----------|-----------------|-----------------|
| Anidulafungin AND 0.002-32 µg/mL | 071425117 | 14.07.2025      | 2028.07.13      |

Ref.

92155 – 921550 – 921551

### Antimicrobial Susceptibility Testing

Tested according to in-house methodology for antifungal agents

| Control strains                            | Medium        | Inoculum                   | Incubation                   | Expected Results<br>MIC range (mg/L) | Results<br>MIC(mg/L) |
|--------------------------------------------|---------------|----------------------------|------------------------------|--------------------------------------|----------------------|
| <i>Candida parapsilosis</i><br>ATCC® 22019 | R.P.M.I. Agar | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>24-48 h | 0.5–4                                | 1                    |
| <i>Candida krusei</i><br>ATCC® 6258        | R.P.M.I. Agar | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>24-48 h | 0.016–0.125                          | 0.094                |
| <i>Candida albicans</i><br>ATCC® 90028     | R.P.M.I. Agar | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>24-48 h | 0.002–0.008                          | 0.008                |

### Batch Release

Approved

Date

18.07.2025

Signature

Quality Control

(This document has been  
established electronically and  
is valid without signature)

The results reported were obtained at the time of release.