Liofilchem®

Certificate of Analysis

Product	Batch	Production date	Expiration date
Vancomycin VA 0.016-256 µg/mL	041025006	10.04.2025	2028.04.09

Ref.

92057 – 920570 – 920571

Antimicrobial Susceptibility Testing
Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
Staphylococcus aureus ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.5–2	0.75
Enterococcus faecalis ATCC® 29212	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–4	2
Streptococcus pneumoniae ATCC® 49619	Mueller Hinton II Agar (Sheep blood 5%)	0.5 McFarland in broth	35 ± 2°C, 5% CO_2 20-24 h	0.12–0.5	0.25

Batch Release

Approved

Date 16.04.2025

Signature Quality Control
(This document has been established electronically and is valid without signature)

The results reported were obtained at the time of release.