



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name: Xpert® HCV Viral Load

Cepheid Catalogue Part No.: GXHCV-VL-CE-10

Kit Lot No.: 1001498818

Cartridge Lot No.: 52301

Kit Expiration Date: 2027-02-07

Legal Manufacturer

Cepheid AB
Röntgenvägen 5
SE-17154 Solna
Sweden

Manufacturing Facility

Cepheid AB
Röntgenvägen 5
SE-171 54 Solna
Sweden

☒ Solna ☐ Sunnyvale

Functional Testing according to D51112 Rev: G

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
LOW Positive	HCV DETECTED: 1.50-1.90 log ₁₀ IU/mL	Passed
HIGH Positive	HCV DETECTED: 4.25-4.55 log ₁₀ IU/mL	Passed
Negative	HCV NOT DETECTED	Passed

☒ If checked this document is produced electronically and valid without a wet signature.

William Palmén

William Palmén (Sep 11, 2025 11:01:08 GMT+2)

Sep 11, 2025

Signature of Quality Assurance,

Date

Name: William Palmén

Title: QA Analyst