

## **CERTIFICATE OF ANALYSIS**

Product Name:	Prolex™ Staph Xtra Latex Kit		Lot Number:	C1249	
Product Code:	PL.1080		Expiration Date:	2026-09-30	
Product Code	Quantity	Compone	nts		
PL.1083	2 x 2.5 ml	Staph Xtra	Latex Reagent		
PL.1085	1 x 5.0 ml	Staph Xtra	a Negative Control Latex Reagent		
	200	Mixing Stic	ks		
	20 cards	Test Cards	<b>:</b>		
Additional Information					
Storage Conditions: 2-8°C					
Instructions for Use: Refer to pro-lab.com/resources for current version					
Quality Control Testing					
Criteria				Result	
Strong agglutination of Staph Xtra Latex Reagent when tested against <i>Staphylococcus aureus</i> and MRSA* including capsular types 5 and 8 within 20 seconds.				PASS	
No agglutination of Staph Xtra Negative Control when tested against test isolates at 20 seconds.				PASS	
* MRSA - (Methicillin-resistant <i>Staphylococcus aureus</i> )					
Release					
2025-02-21			pra		
Date			Quality Control		

## Pro-Lab Diagnostics U.K.

3 Bassendale Road, Bromborough, Wirral,
Merseyside CH62 3QL
United Kingdom
Tel.: 0151 353 1613
www.pro-lab.co.uk
E-mail: uksupport@pro-lab.co.uk

20 Mural Street, Unit #4
Richmond Hill, Ontario, L4B 1K3
Canada

**Pro-Lab Diagnostics Canada** 

Canada
Tel.: 905-731-0300
www.pro-lab.com
E-mail: support@pro-lab.com

## Pro-Lab Diagnostics U.S.A.

1301 Blue Ridge Drive, Suite #101 Georgetown, Texas 78626-3269 United States Tel.: 512-832-9145 www.pro-lab-direct.com E-mail: support@pro-lab.us