

## CERTIFICATE OF ANALYSIS

|                                                                              |                                    |
|------------------------------------------------------------------------------|------------------------------------|
| <b>Product Name:</b> Streptococcal Grouping Reagent Latex Suspension Group F | <b>Lot Number:</b> C1247E          |
| <b>Product Code:</b> PL.035                                                  | <b>Expiration Date:</b> 2026-11-30 |

### Additional Information

Storage Conditions: 2-8°C  
 Instructions for Use: Refer to pro-lab.com/resources for current version

### Quality Control Testing

| Appearance                                            | Reaction | Result       |
|-------------------------------------------------------|----------|--------------|
| Homogenous suspension with no agglutination           | N/A      | Satisfactory |
| <b>Organism</b>                                       |          |              |
| <i>Streptococcus</i> sp. ATCC 12392 (Group F, type 2) | Positive | PASS         |
| <i>Streptococcus</i> Group F - Clinical Isolate       | Positive | PASS         |
| <i>Streptococcus pyogenes</i> ATCC 19615 (Group A)    | Negative | PASS         |
| <i>Streptococcus agalactiae</i> ATCC 12386 (Group B)  | Negative | PASS         |
| <i>Streptococcus equisimilis</i> ATCC 12388 (Group C) | Negative | PASS         |
| <i>Enterococcus faecalis</i> ATCC 49477 (Group D)     | Negative | PASS         |
| <i>Streptococcus</i> sp. ATCC 12394 (Group G)         | Negative | PASS         |
| <i>Listeria monocytogenes</i> NCTC 11994              | Negative | PASS         |
| <b>Control</b>                                        |          |              |
| Polyvalent Positive Control (PL.040)                  | Positive | PASS         |

### Release

|            |                                                                                      |
|------------|--------------------------------------------------------------------------------------|
| 2025-04-25 |  |
| Date       | Quality Control                                                                      |

#### Pro-Lab Diagnostics U.K.

3 Bassendale Road, Bromborough, Wirral,  
 Merseyside CH62 3QL  
 United Kingdom  
 Tel.: 0151 353 1613  
 www.pro-lab.co.uk  
 E-mail: uksupport@pro-lab.co.uk

#### Pro-Lab Diagnostics Canada

20 Mural Street, Unit #4  
 Richmond Hill, Ontario, L4B 1K3  
 Canada  
 Tel.: 905-731-0300  
 www.pro-lab.com  
 E-mail: support@pro-lab.com

#### Pro-Lab Diagnostics U.S.A.

1301 Blue Ridge Drive, Suite #101  
 Georgetown, Texas 78626-3269  
 United States  
 Tel.: 512-832-9145  
 www.pro-lab-direct.com  
 E-mail: support@pro-lab.us