

# Liofilchem®

## Certificate of Analysis

Page 1 of 1

| Product               | Batch     | Expiration date |
|-----------------------|-----------|-----------------|
| Cefiderocol FDC 30 µg | 080625065 | 2026.08.06      |

### Ref.

9266 – 9266/1 – 9266/2

### Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

| Control strains                              | Medium                    | Inoculum                   | Incubation                   | Expected Results<br>Zone range (mm) | Results<br>Zone (mm) |
|----------------------------------------------|---------------------------|----------------------------|------------------------------|-------------------------------------|----------------------|
| <i>Escherichia coli</i><br>ATCC® 25922       | Mueller Hinton II<br>Agar | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>16-20 h | 25-31 (CLSI)<br>24-30 (EUCAST)      | 26                   |
| <i>Pseudomonas aeruginosa</i><br>ATCC® 27853 | Mueller Hinton II<br>Agar | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>16-20 h | 22-31 (CLSI)<br>23-29 (EUCAST)      | 27                   |

### Batch Release

Approved

Date 12.08.2025

Signature

Quality Control  
(This document has been  
established electronically and  
is valid without signature)

The results reported were obtained at the time of release.