

QUALITY CONTROL CERTIFICATE

ESBL + AmpC Screen ID pack, Neo-sensitabs

REF: 98008

LOT: 25-4006

Date of expiry: 2026-04

Appearance: Light yellow, plane tablet, 9 mm in diameter, on both sides printed with a black identification code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:	Results
E. coli ATCC 25922	Negative
K. pneumoniae ATCC700603	ESBL Positive
Ent. cloaceon NCTC 13406	AmpC Positive

Test strains:	Results
E. coli ATCC 25922	Negative
K. pneumoniae ATCC700603	ESBL Positive
Ent. cloaceon NCTC 13406	AmpC Positive

Product complies with specification.

Uniformity: Inhibition zones of individual tablets vary less than + 1 mm from mean
Complies.

Microbiological quality: Complies.

Rosco Diagnostica ApS



Claus Bo Svendsen MD PhD MBA
Chief Medical Officer, Head of Quality