

CERTIFICATE OF ANALYSIS

Product Name: Staph Latex Kit		Lot Number: C1771
Product Code: PL.080B		Expiration Date: 2026-12-31
Product Code	Quantity	Components
PL.083B	2 x 2.5 ml	Staph Latex Reagent
PL.085B	1 x 2.5 ml	Negative Control Latex Reagent
	200	Mixing Sticks
PL.092-20	20 cards	Test Cards
Additional Information		
Storage Conditions:		2-8°C
Instructions for Use:		Refer to pro-lab.com/resources for current version
Quality Control Testing		
Criteria		Result
Strong agglutination of Staph Latex Reagent when tested against <i>Staphylococcus aureus</i> .		PASS
No agglutination of Negative Control when tested against test isolates.		PASS
Release		
2025-09-09		
Date	Quality Control	

Pro-Lab Diagnostics U.K.
3 Bassendale Road, Bromborough, Wirral,
Merseyside CH62 3QL
United Kingdom
Tel.: 0151 353 1613
www.pro-lab.co.uk
E-mail: uksupport@pro-lab.co.uk

Pro-Lab Diagnostics Canada
20 Mural Street, Unit #4
Richmond Hill, Ontario, L4B 1K3
Canada
Tel.: 905-731-0300
www.pro-lab.com
E-mail: support@pro-lab.com

Pro-Lab Diagnostics U.S.A.
1301 Blue Ridge Drive, Suite #101
Georgetown, Texas 78626-3269
United States
Tel.: 512-832-9145
www.pro-lab-direct.com
E-mail: support@pro-lab.us