



CERTIFICATE OF ANALYSIS

SI 390.900 – URO-QUICK SCREENING KIT

Code

MIC_COA_SI390900#SCR250398

PRODUCT NAME	URO-QUICK SCREENING KIT
REF NUMBER	SI 390.900
LOT	SCR250398
EXPIRATION DATE (yyyy-mm-dd)	2027-02-14

Tests	Specifications	Results
Visual inspection		
Appearance of the liquid culture medium Appearance of the vial	Limpid, yellow, transparent liquid Green ring cap, pierceable septum, white stripe, presence of the stirrer, absence of defects	OK
Microbiological activity		
Escherichia coli ATCC 8739	Growth detection with inoculum ≤ 100 CFU ¹	OK
Microbiological state		
72 hours of incubation at 37°C	Absence of turbidity of the liquid culture media ²	OK

¹Inoculum concentration guaranteed using certified standards.

²Visual inspection of the liquid culture medium after incubation

The sampling method and quality controls have been performed in accordance with the procedures in force.

Analysis outcome:	<input checked="" type="checkbox"/> Compliant <input type="checkbox"/> Not compliant
-------------------	---

Lot approved on	2026-01-07	By: (Quality Assurance)	<i>Gigante Micaela</i>
-----------------	------------	----------------------------	------------------------

©Copyright [2025]. This document is exclusive property of ALIFAX S.r.l., confidential and not freely disclosable. All rights reserved.

ALIFAX S.r.l.

Registered and head office
via Petrarca, 2/1 - 35020 Polverara PD Italy
Tel. (+39) 049 0992000 Fax (+39) 049 5855434
info@alifax.com - info.alifax@pec.alifax.com

R&D, Design and Manufacturing Dept
via Merano, 30,33045 Nimis UD Italy
Tel. (+39) 0432 197900 Fax (+39) 0432 547378
info.udine@alifax.com - info.udine@pec.alifax.com

Entered in Padova companies register at the n. 04337640280. Company stock 10.000.000 entirely deposited.
Company with single partner and subjected to direction and coordination of Alifax Holding S.p.A.

www.alifax.com