

Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

Product	Anti-Coli II
Catalogue no.	TR2121-01
Package size	5 ml
Lot no.	7961022
Expiry date	2027-09
Intended use	Anti-Coli II records all <i>E. coli</i> strains, as long as they contain the antigens O 55:(K 59), O 86:(K 61), O 91:(K-), O 111:(K 58), O 119:(K 69), O 126:(K 71), O127:(K 63) and O 128:(K 67).
Specificity	No agglutination with <i>E. coli</i> strains which are recorded by Anti-Coli I and Anti-Coli III.
Preservatives	Sodium azide 0.9 mg/ ml
Storage	2 °C...8 °C

Berlin,
place, date

13.12.2022


signature Head of Quality Control