



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® C. difficile BT

Catalogue Part No.: GXCDIFFBT-CE-10

Kit Lot No.: 1001510848

Cartridge Lot No.: 38801

Kit Expiration Date: 2027-05-09

Legal Manufacturer

Cepheid AB
Röntgenvägen 5
SE-171 54 Solna,
Sweden

Manufacturing Facility

Cepheid AB
Röntgenvägen 5
SE-171 54 Solna,
Sweden

Functional Testing

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
Positive Control	Toxigenic C.diff POS;Binary Toxin POS;027 PRESUMPTIVE POS	Passed
Negative Control	Toxigenic C.diff NEG;Binary Toxin NEG;027 NEG	Passed

If checked this document is produced electronically and valid without a wet signature.

Alexander Avramidis

Alexander Avramidis (11/27/2025 10:59:33 GMT+1)

Signature of Quality Assurance

Date

Name: Alexander Avramidis

Title: QA Analyst