

# Liofilchem®

## Certificate of Analysis

Page 1 of 1

| Product                         | Batch     | Production date | Expiration date |
|---------------------------------|-----------|-----------------|-----------------|
| FLUCONAZOLE FLU 0.016-256 µg/mL | 052325022 | 23.05.2025      | 2028.05.22      |

### Ref.

92147 – 921470 – 921471

### Antimicrobial Susceptibility Testing

Tested according to in-house methodology for antifungal agents

| Control strains                            | Medium        | Inoculum                   | Incubation                   | Expected Results<br>MIC range (mg/L) | Results<br>MIC(mg/L) |
|--|---------------|----------------------------|------------------------------|--------------------------------------|----------------------|
| <i>Candida krusei</i><br>ATCC® 6258        | R.P.M.I. Agar | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>24-48 h | 128–≥256                             | 256                  |
| <i>Candida parapsilosis</i><br>ATCC® 22019 | R.P.M.I. Agar | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>24-48 h | 1–8                                  | 3                    |
| <i>Candida albicans</i><br>ATCC® 90028     | R.P.M.I. Agar | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>24-48 h | 0.125–0.5                            | 0.25                 |

### Batch Release

Approved

Date

03.06.2025

Signature

Quality Control  
(This document has been  
established electronically and  
is valid without signature)

The results reported were obtained at the time of release.