

## Certificate of Analysis

PRODUCT DESCRIPTION: nal von minden Drug-Screen® KRA 100 Test

PRODUCT	LOT	LOT AMOUNT	EXPIRATION
CODE	NUMBER		DATE
102099	KRA25100003	2010PCS	2027-10-16

### PRODUCT TEST RESULTS

QC Items		Specification	Test Result	Conclusion
Physical	Appearance	Intact	Intact	PASS
Functional Performance	KRA 100 ng/ml	+50%	Positive (+)	PASS
		-50%	Negative (-)	PASS
		3 x Cut Off	Positive (+)	PASS
		Negative Sample	Negative (-)	PASS

**Conclusion:** This batch of product meets the QC specification

Accepted:



Miriam Lang  
 Director QM / RA  
 Quality Management & Regulatory  
 Affairs



Regensburg, 2025-11-17

We help you care!