

## • Certificate of Analysis

Product Name : **IVD Matrix HCCA-portioned**

Part No.: **8290200**

LOT No.: **6020225117**

Use by<sup>1)</sup>: **2027-11-03**

Date of Production: **2025-11-04**

Storage Conditions : **Store in a refrigerator at 2°C to 8°C on arrival.**

| <u>Test</u>                                                                   | <u>Specification</u>                                                                                              | <u>LOT Results</u> |
|-------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------|--------------------|
| <b>1. Amount of HCCA content</b>                                              | AQL-value=1.0<br>Inspection level II<br>U=2.8 mg // L=2.2 mg<br>Sampling in compliance of DIN ISO 3951-1          | Passed             |
| <b>2. MALDI Biotyper identification of a 1:20 diluted IVD BTS</b>             | Classification Result Report lists " <i>Escherichia coli</i> " as best match with a log score value $\geq 2.00$ . | Passed             |
| <b>3. Blank value analysis of IVD Matrix HCCA-portioned by MALDI Biotyper</b> | Classification Result Report lists "No Organism Identification Possible." with a log score value $< 1.70$ .       | Passed             |

### Remarks:

1) IVD Matrix HCCA-portioned is shipped at ambient temperature. The expiry date on the package is valid for the product when stored in a refrigerator at 2°C to 8°C on arrival.

Bruker Daltonics GmbH & Co. KG  
Fahrenheitstr. 4  
28359 Bremen  
Germany  
Phone +49-421-2205-0

QC release date: 2025-11-27

  
Quality Management