

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Neomycin N 30 µg	040325044	2028.04.02

Ref.

9032 – 9032/1 – 9032/2

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	14-20*	20
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	16-22	21

*Established and validated by EUCAST

Batch Release

Approved

Date

14.04.2025

Signature

Quality Control
(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.