


### CERTIFICATE OF ANALYSIS

<b>Product Name: Staph Latex Kit</b>		<b>Lot Number: C2193</b>
<b>Product Code: PL.080B</b>		<b>Expiration Date: 2027-05-31</b>
<b>Product Code</b>	<b>Quantity</b>	<b>Components</b>
PL.083B	2 x 2.5 ml	Staph Latex Reagent
PL.085B	1 x 2.5 ml	Negative Control Latex Reagent
	200	Mixing Sticks
PL.092-20	20 cards	Test Cards
<b>Additional Information</b>		
Storage Conditions:		2-8°C
Instructions for Use:		Refer to pro-lab.com/resources for current version
<b>Quality Control Testing</b>		
<b>Criteria</b>		<b>Result</b>
Strong agglutination of Staph Latex Reagent when tested against <i>Staphylococcus aureus</i> .		PASS
No agglutination of Negative Control when tested against test isolates.		PASS
<b>Release</b>		
2026-01-02		
Date	Quality Control	

**Pro-Lab Diagnostics U.K.**  
3 Bassendale Road, Bromborough, Wirral,  
Merseyside CH62 3QL  
United Kingdom  
Tel.: 0151 353 1613  
www.pro-lab.co.uk  
E-mail: uk-support@pro-lab.co.uk

**Pro-Lab Diagnostics Canada**  
20 Mural Street, Unit #4  
Richmond Hill, Ontario, L4B 1K3  
Canada  
Tel.: 905-731-0300  
www.pro-lab.com  
E-mail: support@pro-lab.com

**Pro-Lab Diagnostics U.S.A.**  
1301 Blue Ridge Drive, Suite #101  
Georgetown, Texas 78626-3269  
United States  
Tel.: 512-832-9145  
www.pro-lab-direct.com  
E-mail: support@pro-lab.us