

### CERTIFICATE OF ANALYSIS

<b>Product Name:</b> Streptococcal Grouping Reagent Latex Suspension Group G	<b>Lot Number:</b> C1817B	
<b>Product Code:</b> PL.036	<b>Expiration Date:</b> 2027-05-31	

#### Additional Information

Storage Conditions:	2-8°C
Instructions for Use:	Refer to pro-lab.com/resources for current version

#### Quality Control Testing

Appearance	Reaction	Result
Homogenous suspension with no agglutination	N/A	Satisfactory
<b>Organisms</b>		
<i>Streptococcus</i> sp. ATCC 12394 (Group G)	Positive	PASS
<i>Streptococcus</i> Group G - Clinical Isolate	Positive	PASS
<i>Streptococcus pyogenes</i> ATCC 19615 (Group A)	Negative	PASS
<i>Streptococcus agalactiae</i> ATCC 12386 (Group B)	Negative	PASS
<i>Streptococcus equisimilis</i> ATCC 12388 (Group C)	Negative	PASS
<i>Enterococcus faecalis</i> ATCC 49477 (Group D)	*	PASS
<i>Streptococcus</i> sp. ATCC 12392 (Group F, type 2)	Negative	PASS
<i>Enterococcus faecalis</i> ATCC 19433	Negative	PASS
<i>Enterococcus faecalis</i> ATCC 49478	Negative	PASS
<i>Enterococcus faecalis</i> ATCC 29212	Negative	PASS
<i>Listeria monocytogenes</i> NCTC 11994	Negative	PASS
<b>Control</b>		
Polyvalent Positive Control (PL.040)	Positive	PASS

\*Cross reaction may be present.

#### Release

2025-09-30	
Date	Quality Control

**Pro-Lab Diagnostics U.K.**  
3 Bassendale Road, Bromborough, Wirral,  
Merseyside CH62 3QL  
United Kingdom  
Tel.: 0151 353 1613  
www.pro-lab.co.uk  
E-mail: uksupport@pro-lab.co.uk

**Pro-Lab Diagnostics Canada**  
20 Mural Street, Unit #4  
Richmond Hill, Ontario, L4B 1K3  
Canada  
Tel.: 905-731-0300  
www.pro-lab.com  
E-mail: support@pro-lab.com

**Pro-Lab Diagnostics U.S.A.**  
1301 Blue Ridge Drive, Suite #101  
Georgetown, Texas 78626-3269  
United States  
Tel.: 512-832-9145  
www.pro-lab-direct.com  
E-mail: support@pro-lab.us