

• Certificate of Analysis

Product Name : **IVD Bacterial Test Standard**

Part No.: 8290190

LOT No.: 6030325020

Use by¹⁾: 2027-12-08

Date of Production: 2025-12-08

Storage Conditions : Store at -18 °C or below immediately on arrival.

| <u>Test</u> | <u>Specification</u> | <u>LOT Results</u> |
|---|---|--------------------|
| Verification of the calibration points by MALDI-TOF MS | Within a mass range of ± 900 ppm | Passed |
| Relative intensity of reference protein No. 3 by MALDI-TOF MS | 0.20 ± 0.05 | Passed |
| Relative intensity of reference protein No. 4 by MALDI-TOF MS | 0.22 ± 0.05 | Passed |
| MALDI Biotyper identification | Classification Result Report lists " <i>Escherichia coli</i> " as best match. | Passed |
| | log(score) Value ≥ 2.2 of the best match | Passed |

Remarks:

¹⁾ IVD BTS is shipped at ambient temperature but must be stored at -18°C or below immediately on arrival. The expiry date on the package is valid for the enclosed IVD BTS when stored at -18°C or below on arrival. Frozen, dissolved Bruker IVD BTS can be stored for up to 5 months at -18°C or below. See instruction for use at www.bruker.com/care.

Bruker Daltonics GmbH & Co. KG
 Fahrenheitstr. 4
 28359 Bremen
 Germany
 Phone +49-421-2205-0


 Quality Management

QC release date: 2025-12-18