



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® MRSA/SA Blood Culture

Cepheid Catalogue Part No.: GXM RSA/SABC-CE-10

Kit Lot No.: 1001510338

Cartridge Lot No.: 23204

Kit Expiration Date: 2027-08-29

Legal Manufacturer

Cepheid AB
Röntgenvägen 5
SE-17154 Solna
Sweden

Manufacturing Facility

Cepheid
121 N Guild Avenue
Lodi, CA 95240
USA

Solna

Sunnyvale

Lodi

Functional Testing according to D36872, Rev. H

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
Positive	MRSA Positive; SA Positive	Passed
Negative	MRSA Negative; SA Negative	Passed

If checked, this document is produced electronically and therefore valid without a wet signature



Signature of Quality Assurance,

23 NOV 25

Date

Name: Chong Moua

Title: Quality Assurance Specialist