

# Liofilchem®

## Certificate of Analysis

Product	Batch	Expiration date
<b>Piperacillin-tazobactam TZP 0.016/4-256/4</b>	030226023	2029.03.01

**Ref.**

92108 – 921080 – 921081

**Antimicrobial Susceptibility Testing**

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (µg/mL)	Results MIC (µg/mL)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–8 <sup>a,b</sup>	2
<i>Pseudomonas aeruginosa</i> ATCC® 27853	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–8 <sup>a,b</sup>	4
<i>Escherichia coli</i> ATCC® 35218	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.5–2 <sup>a,b</sup>	1
<i>Klebsiella pneumoniae</i> ATCC® 700603	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	8–32 <sup>a,b</sup>	16

a) CLSI M100-Ed35

b) EUCAST Quality Control v.16.0

**Batch Release**

Approved

**Date**

04.03.2026

**Signature**Quality Control  
(This document has been  
established electronically and  
is valid without signature)

The results reported were obtained at the time of release.