

Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

sifin diagnostics gmbh

Berliner Allee 317-321
13088 Berlin, Germany

T: +49 30 700 144-0
F: +49 30 700 144-30
info@sifin.de
www.sifin.de

Product	Anti-Coli O 128:K 67
Catalogue no.	TS2214
Package size	1 ml
Lot no.	0290125
Expiry date	2029-12
Intended use	Anti-Coli O 128:K 67 records all <i>E. coli</i> strains that contain the antigen O 128:K 67 in slide agglutination. In the Widal reaction the serum reaches a titre of 1:1600 (tube and MTP) with the homologous <i>E. coli</i> strains; deviations of ± 1 titre step are acceptable.
Specificity	Since unspecific co-agglutinations with other <i>E. coli</i> strains often occur in the slide agglutination test, a positive result in the slide test must always be confirmed by examination of the boiled <i>E. coli</i> isolate in the Widal test.
Preservatives	Sodium azide 0.9 mg/ ml
Storage	2 °C...8 °C

This document has been established electronically and is valid without signature.
The batch release was done in accordance with SOP QMV 7.