

### CERTIFICATE OF ANALYSIS

|  |                                    |  |
|--|------------------------------------|--|
| <b>Product Name:</b> Streptococcal Grouping Reagent Latex Suspension Group C | <b>Lot Number:</b> C1814E          |  |
| <b>Product Code:</b> PL.033  | <b>Expiration Date:</b> 2027-05-31 |  |


#### Additional Information

|                       |  |
|-----------------------|--|
| Storage Conditions:   | 2-8°C  |
| Instructions for Use: | Refer to pro-lab.com/resources for current version |

#### Quality Control Testing

| Appearance  | Reaction | Result       |
|---|----------|--------------|
| Homogenous suspension with no agglutination           | N/A      | Satisfactory |
| <b>Organisms</b>                                      |          |              |
| <i>Streptococcus equisimilis</i> ATCC 12388 (Group C) | Positive | PASS         |
| <i>Streptococcus</i> Group C - Clinical Isolate       | Positive | PASS         |
| <i>Streptococcus pyogenes</i> ATCC 19615 (Group A)    | Negative | PASS         |
| <i>Streptococcus agalactiae</i> ATCC 12386 (Group B)  | Negative | PASS         |
| <i>Enterococcus faecalis</i> ATCC 49477 (Group D)     | Negative | PASS         |
| <i>Streptococcus</i> sp. ATCC 12392 (Group F, type 2) | Negative | PASS         |
| <i>Streptococcus</i> sp. ATCC 12394 (Group G)         | Negative | PASS         |
| <i>Streptococcus</i> Group D - Clinical Isolate       | Negative | PASS         |
| <i>Listeria monocytogenes</i> NCTC 11994              | Negative | PASS         |
| <b>Control</b>  |          |              |
| Polyvalent Positive Control (PL.040)                  | Positive | PASS         |

#### Release

|            |  |
|------------|--|
| 2025-11-05 |  |
| Date       | Quality Control  |

**Pro-Lab Diagnostics U.K.**  
3 Bassendale Road, Bromborough, Wirral,  
Merseyside CH62 3QL  
United Kingdom  
Tel.: 0151 353 1613  
www.pro-lab.co.uk  
E-mail: uksupport@pro-lab.co.uk

**Pro-Lab Diagnostics Canada**  
20 Mural Street, Unit #4  
Richmond Hill, Ontario, L4B 1K3  
Canada  
Tel.: 905-731-0300  
www.pro-lab.com  
E-mail: support@pro-lab.com

**Pro-Lab Diagnostics U.S.A.**  
1301 Blue Ridge Drive, Suite #101  
Georgetown, Texas 78626-3269  
United States  
Tel.: 512-832-9145  
www.pro-lab-direct.com  
E-mail: support@pro-lab.us