

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
Posaconazole POS 0.002-32 µg/mL	012126047	21.01.2026	2029.01.20

Ref.

92152 – 921520 – 921521

Antimicrobial Susceptibility Testing

Tested according to in-house methodology for antifungal agents

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Candida krusei</i> ATCC® 6258	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.125-0.5	0.125
<i>Candida parapsilosis</i> ATCC® 22019	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.032-0.25	0.032
<i>Candida albicans</i> ATCC® 90028	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.032-0.125	0.047

Batch Release

Approved

Date

26.01.2026

Signature

Quality Control
(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.