

# Liofilchem®

## Certificate of Analysis

Page 1 of 1

| Product                       | Batch     | Production date | Expiration date |
|-------------------------------|-----------|-----------------|-----------------|
| Vancomycin VA 0.016-256 µg/mL | 041326041 | 13.04.2026      | 2029.04.12      |

Ref.

92057 – 920570 – 920571

### Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

| Control strains                                | Medium                                        | Inoculum                   | Incubation                              | Expected Results<br>MIC range (mg/L) | Results<br>MIC(mg/L) |
|------------------------------------------------|-----------------------------------------------|----------------------------|-----------------------------------------|--------------------------------------|----------------------|
| <i>Staphylococcus aureus</i><br>ATCC® 29213    | Mueller Hinton II<br>Agar                     | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>16-20 h            | 0.5–2                                | 0.75                 |
| <i>Enterococcus faecalis</i><br>ATCC® 29212    | Mueller Hinton II<br>Agar                     | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>16-20 h            | 1–4                                  | 2                    |
| <i>Streptococcus pneumoniae</i><br>ATCC® 49619 | Mueller Hinton II<br>Agar (Sheep blood<br>5%) | 0.5 McFarland in<br>broth  | 35 ± 2°C, 5% CO <sub>2</sub><br>20-24 h | 0.12–0.5                             | 0.25                 |

### Batch Release

Approved

Date

20.04.2026

Signature

Quality Control

(This document has been  
established electronically and  
is valid without signature)

The results reported were obtained at the time of release.