

Quality Control

Certificate of Analysis

QC release date: 2026-05-18

Product: Glucose Hexokinase FS
Product Code: 12511

Cat. No.: 1 2511 99 10 021

Delivery of kits:	No of kits:	1
	Kit Lot/Los:	60179458
	Kit Expiry:	2028-05-07

Delivery of bulk:	Volume:
	Processing Lot:

Results for:

Reagent 1	Lot No.:	39051
	Expiry:	2028-05-07

Reagent 2	Lot No.:	39051
	Expiry:	2028-05-07

Examination:

Appearance:	Reagent 1:	clear, colourless
	Reagent 2:	clear, colourless

Odour:	Reagent 1:	no characteristic odour
	Reagent 2:	no characteristic odour

pH:	Reagent 1 (25 °C ± 1 °C):	7.79
	Reagent 2 (25 °C ± 1 °C):	7.02
	R1+R2 (25 °C ± 1 °C):	7.76

Absorbance:	Reagent 1:	0.109
340 nm / 15-25 °C	R1 + R2:	0.090

Control sera:	TruLab N	TruLab P
	Lot: 36871	Lot: 36872

Fresh (at 37°C)	92.3 mg/dL	267 mg/dL
-----------------	------------	-----------

Target values	92.7 77.9 - 108	271 228 - 315
---------------	-----------------	---------------

Additional remarks:

This lot meets our specifications.

This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. D. Thönges

DiaSys Diagnostic Systems GmbH
65558 Holzheim, GERMANY

Quality Control

Certificate of Analysis

QC release date: 2026-05-18

Product: GL Control N
Product Code: 920536

Cat. No.: 920536

Delivery of kits:	No of kits:	1
	Kit Lot/Los:	1_44067
	Kit Expiry:	2027-09-30

Results for:

Control	Lot No.:	44067
	Expiry:	2027-09-30

Assigned values see insert.

Additional remarks:

This lot meets our specifications.
This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. D. Thönges

DiaSys Diagnostic Systems GmbH
65558 Holzheim, GERMANY

Quality Control

Certificate of Analysis

QC release date: 2026-03-25

Product: GL Control P
Product Code: 920537

Cat. No.: 920537

Delivery of kits:	No of kits:	1
	Kit Lot/Los:	3_43861
	Kit Expiry:	2027-07-31

Results for:

Control	Lot No.:	43861
	Expiry:	2027-07-31

Assigned values see insert.

Additional remarks:

This lot meets our specifications.
This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. D. Thönges

DiaSys Diagnostic Systems GmbH
65558 Holzheim, GERMANY

Quality Control

Certificate of Analysis

QC release date: 2026-06-19

Product: System Solution SensoStar
Product Code: 920532

Cat. No.: 920532

Delivery of kits:	No of kits:	7
	Kit Lot/Los:	2_44214
	Kit Expiry:	2027-12-31

Results for:

Reagent	Lot No.:	44214
	Expiry:	2027-12-31

Examination:

Conductivity (25 °C)	5.55 mS/cm
------------------------	------------

Additional remarks:

This lot meets our specifications.
This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. D. Thönges