

Quality Control

Certificate of Analysis

QC release date: 2026-05-07

Product: Lp(a) 21 FS
Product Code: 17139

Cat. No.: 1 7139 99 10 931

Delivery of kits:	No of kits:	6
	Kit Lot/Los:	60179204
	Kit Expiry:	2028-04-30

Delivery of bulk:	Volume:
	Processing Lot:

Results for:

Reagent 1	Lot No.:	39162
	Expiry:	2028-04-30

Reagent 2	Lot No.:	39162
	Expiry:	2028-04-30

Examination:

Appearance:	Reagent 1:	clear to slightly opalescent, yellowish
-------------	------------	---

Reagent 2:	milky, white
------------	--------------

Odour:	Reagent 1:	no characteristic odour
	Reagent 2:	no characteristic odour

Control sera:	TruLab Lp(a) 1	TruLab Lp(a) 2
	Lot: 38312	Lot: 38313

Test results	43.6 nmol/L	112 nmol/L
--------------	-------------	------------

Target values	42.4 33.9 - 50.9	114 91.2 - 137
---------------	---------------------	-------------------

Additional remarks:

This lot meets our specifications.
This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. D. Thönges