



EU Quality Management System Certificate

Regulation (EU) 2017/746 on in Vitro Diagnostic Medical Devices, Annex IX Chapter I

Certificate No. V13 064137 0051 Rev. 00

Manufacturer:



Bruker Daltonics GmbH & Co. KG

Fahrenheitstrasse 4
28359 Bremen
GERMANY

SRN Manufacturer - DE-MF-000024392

The quality management system has been evaluated in accordance with Regulation (EU) 2017/746, Annex IX Chapter I with a positive result.

Details on devices covered by the quality management system are described on the following page(s). The report referenced below summarises the results of the assessment and includes reference to relevant CS, harmonised standards and test reports.

The certified quality management system is subject to periodical surveillance.

If class A devices in sterile conditions are covered by this certificate, the audit was limited to the aspects relating to establishing, securing, and maintaining sterile conditions.

If class B or C excluding self-/near-patient-testing, or class C companion diagnostics devices are covered by this certificate, the quality management system assessment was accompanied by the assessment of technical documentation for devices selected on a representative basis. The periodical surveillance includes further assessment of the technical documentation on the basis of representative samples.

If class D devices, class B or C self-/near-patient testing, or class C companion diagnostics devices are covered by this certificate, an EU Technical Documentation Assessment Certificate in accordance with Annex IX Chapter II is required before placing them on the market.

All applicable requirements of the Testing, Certification, Validation and Verification Regulations TÜV SÜD Group have to be complied with.

For details and certificate validity see: www.tuvsud.com/ps-cert?q=cert:V13 064137 0051 Rev. 00

Report No.: 713352719, 713352719_CN

Preceding Certificate No.: V12 064137 0048 Rev. 01

Valid from: 2025-05-06

Valid until: 2028-12-17

Marta Carnielli
Head of Certification IVD

Issue date: 2025-05-06



EU Quality Management System Certificate

Regulation (EU) 2017/746 on in Vitro Diagnostic Medical Devices, Annex IX Chapter I

Certificate No. V13 064137 0051 Rev. 00

Classification: Class C
Device Group: W0203 + IVP 3012 - Microbiology (Culture) Instruments
Intended Purpose: IVD Medical device Software for Microbiology (Culture)

Classification: Class C
Device Group: W0104 + IVP 3013 - Microbiology (Culture)
Intended Purpose: IVD Reagents for Microbiology (Culture)

Classification: Class C
Device Group: W0104 + IVP 3012 - Microbiology (Culture)
Intended Purpose: IVD Reagents for Microbiology (Culture)

The validity of this certificate depends on conditions and/or is limited to the following: -none-

Rev.	Dated	Report	Description
00	2025-05-06	713352719, 713352719_C N	Supplemented: Device(s)/group of device(s) added Administrative merge / transfer to new Certificate Type