

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Clotrimazole CLO 50 µg	050625195	2028.05.05

Ref.

9097 – 9097/1 – 9097/2

Antimicrobial Susceptibility Testing

Tested according to in-house disc diffusion method for antifungal agents

Control strains	Medium	Inoculum	Incubation	Results Halo (mm)
<i>Candida albicans</i> ATCC® 90028	Mueller Hinton+ Glucose+ Methylen Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	32
<i>Candida parapsilosis</i> ATCC® 22019	Mueller Hinton+ Glucose+ Methylen Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	31
<i>Candida krusei</i> ATCC® 6258	Mueller Hinton+ Glucose+ Methylen Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	34
<i>Candida tropicalis</i> ATCC® 750	Mueller Hinton+ Glucose+ Methylen Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	26

Batch Release

Approved

Date 14.05.2025

Signature

Quality Control
(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.