

## CERTIFICATE OF ANALYSIS

|                                                                    |                                    |
|--------------------------------------------------------------------|------------------------------------|
| <b>Product Name:</b> Streptococcal Grouping - Extraction Reagent 3 | <b>Lot Number:</b> C2118A          |
| <b>Product Code:</b> PL.039                                        | <b>Expiration Date:</b> 2027-12-31 |

### Additional Information

|                       |                                                    |
|-----------------------|----------------------------------------------------|
| Storage Conditions:   | 2-8°C                                              |
| Instructions for Use: | Refer to pro-lab.com/resources for current version |

### Quality Control Testing

| Organism                                              | Reaction* | Result |
|-------------------------------------------------------|-----------|--------|
| <i>Streptococcus pyogenes</i> ATCC 19615 (Group A)    | Positive  | PASS   |
| <i>Streptococcus agalactiae</i> ATCC 12386 (Group B)  | Positive  | PASS   |
| <i>Streptococcus equisimilis</i> ATCC 12388 (Group C) | Positive  | PASS   |
| <i>Enterococcus faecalis</i> ATCC 49477 (Group D)     | Positive  | PASS   |
| <i>Streptococcus</i> sp. ATCC 12392 (Group F)         | Positive  | PASS   |
| <i>Streptococcus</i> sp. ATCC 12394 (Group G)         | Positive  | PASS   |

\*When used in conjunction with Extraction Reagents 1 and 2 to extract group specific carbohydrate antigens and tested against homologous latex reagents.

### Release

|            |                                                                                      |
|------------|--------------------------------------------------------------------------------------|
| 2026-02-20 |  |
| Date       | Quality Control                                                                      |

#### Pro-Lab Diagnostics U.K.

3 Bassendale Road, Bromborough, Wirral,  
Merseyside CH62 3QL  
United Kingdom  
Tel.: 0151 353 1613  
www.pro-lab.co.uk  
E-mail: uksupport@pro-lab.co.uk

#### Pro-Lab Diagnostics Canada

20 Mural Street, Unit #4  
Richmond Hill, Ontario, L4B 1K3  
Canada  
Tel.: 905-731-0300  
www.pro-lab.com  
E-mail: support@pro-lab.com

#### Pro-Lab Diagnostics U.S.A.

1301 Blue Ridge Drive, Suite #101  
Georgetown, Texas 78626-3269  
United States  
Tel.: 512-832-9145  
www.pro-lab-direct.com  
E-mail: support@pro-lab.us