

CERTIFICATE OF ANALYSIS

Product Name: Prolex™ Streptococcal Grouping Latex Kit		Lot Number: C2259
Product Code: PL.030		Expiration Date: 2027-06-30
Product Code	Quantity	Components
PL.031	1 x 3.0 ml	Streptococcal Group A Latex Reagent
PL.032	1 x 3.0 ml	Streptococcal Group B Latex Reagent
PL.033	1 x 3.0 ml	Streptococcal Group C Latex Reagent
PL.034	1 x 3.0 ml	Streptococcal Group D Latex Reagent
PL.035	1 x 3.0 ml	Streptococcal Group F Latex Reagent
PL.036	1 x 3.0 ml	Streptococcal Group G Latex Reagent
PL.037	1 x 3.2 ml	Streptococcal Extraction Reagent 1
PL.038	1 x 3.2 ml	Streptococcal Extraction Reagent 2
PL.039	2 x 8.0 ml	Streptococcal Extraction Reagent 3
PL.040	1 x 2.0 ml	Polyvalent Positive Control (A,B,C,D,F,G)
	300	Mixing Sticks
PL.092-48	48 cards	Test Cards
Additional Information		
Storage Conditions:		2-8°C
Instructions for Use:		Refer to pro-lab.com/resources for current version
Quality Control Testing		
Criteria		Result
Chemical extraction of Strep group specific antigens.		PASS
Strong agglutination of the test isolate when tested against the corresponding latex reagent.		PASS
No agglutination reaction of the test isolate when tested against heterologous latex reagents.		PASS
Strong agglutination of the Positive Control when tested against the latex reagents.		PASS
Release		
2026-01-27		
Date		Quality Control

Pro-Lab Diagnostics U.K.
 3 Bassendale Road, Bromborough, Wirral,
 Merseyside CH62 3QL
 United Kingdom
 Tel.: 0151 353 1613
 www.pro-lab.co.uk
 E-mail: uksupport@pro-lab.co.uk

Pro-Lab Diagnostics Canada
 20 Mural Street, Unit #4
 Richmond Hill, Ontario, L4B 1K3
 Canada
 Tel.: 905-731-0300
 www.pro-lab.com
 E-mail: support@pro-lab.com

Pro-Lab Diagnostics U.S.A.
 1301 Blue Ridge Drive, Suite #101
 Georgetown, Texas 78626-3269
 United States
 Tel.: 512- 832-9145
 www.pro-lab-direct.com
 E-mail: support@pro-lab.us