

Certificate of Analysis

nerbe plus GmbH & Co. KG hereby certifies that the products:

Article number	06-693-5300	06-662-5300
Lot number	29763/2526_04	03040/2627_06
Expiry date	04/2031	06/2031

are manufactured under strictly defined and monitored conditions as per product specifications. This here by certifies that the supplied product complies with following standards of production.

Quality System Compliance: The product conforms to written material specifications and was manufactured under the ISO 9001/2015 quality system.

nerbe plus is certified according to EN ISO 13485.

Material Polypropylene		
Test	Limit	Test Result
Humane DNA	< 2 pg; of template copy	Product is free from Human DNA
Endotoxin	≤ 0.03 EU/ ml	Product is free from Endotoxin
DNase	≤ 1.0 X 10 ⁻⁴ U / μl	Product is free from DNase
RNase	≤ 1.0 X 10 ⁻⁸ U / μl	Product is free from RNase
PCR Inhibitor	Less than 2 CT variation	Product is free of PCR Inhibitor

10.09.2026

Quality Management

nerbe plus GmbH & Co. KG

This certificate was issued digitally and is valid without a signature.