Liofilchem®

Certificate of Analysis

Product	Batch	Expiration date
Piperacillin-tazobactam TZP 0.016/4-256/4	090326021	2029.09.02

Ref.

92108 – 921080 – 921081

Antimicrobial Susceptibility Testing
Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (µg/mL)	Results MIC (µg/mL)
Escherichia coli ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–8 [a,b]	2
Pseudomonas aeruginosa ATCC® 27853	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–8 [a,b]	4
Escherichia coli ATCC® 35218	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.5–2 [a,b]	1
Klebsiella pneumoniae ATCC® 700603	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	8–32 [a,b]	24

a) CLSI M100-Ed35
b) EUCAST Quality Control v.16.0

Batch Release

Approved

Date 11.09.2026

Signature Quality Control
(This document has been established electronically and is valid without signature)

The results reported were obtained at the time of release.