

# Liofilchem®

## Certificate of Analysis

Page 1 of 1

| Product                      | Batch     | Production date | Expiration date |
|------------------------------|-----------|-----------------|-----------------|
| Oxacillin OX 0.016-256 µg/mL | 052026038 | 20.05.2026      | 2029.05.19      |

### Ref.

92015 – 920150 – 920151

### Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

| Control strains                             | Medium                    | Inoculum                   | Incubation                   | Expected Results<br>MIC range (mg/L) | Results<br>MIC(mg/L) |
|---------------------------------------------|---------------------------|----------------------------|------------------------------|--------------------------------------|----------------------|
| <i>Staphylococcus aureus</i><br>ATCC® 29213 | Mueller Hinton II<br>Agar | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>16-20 h | 0.12–0.5                             | 0.19                 |
| <i>Enterococcus faecalis</i><br>ATCC® 29212 | Mueller Hinton II<br>Agar | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>16-20 h | 8–32                                 | 12                   |

### Batch Release

Approved

Date

26.05.2026

Signature

Quality Control  
(This document has been  
established electronically and  
is valid without signature)

The results reported were obtained at the time of release.