Certificate of Analysis

Product Name :	**IVD Matrix HCCA-portioned**
Part No.:	**8290200**
LOT No.:	6020226058
Use by[1]:	2028-04-12
Date of Production:	2026-04-13
Storage Conditions :	**Store in a refrigerator at 2°C to 8°C on arrival.**

Test	Specification	LOT Results

1. Amount of HCCA content	AQL-value=1.0
	Inspection level II
	U=2.8 mg // L=2.2 mg
	Sampling in compliance of DIN ISO 3951-1
		Passed

2. MALDI Biotyper identification of a 1:20 diluted IVD BTS	Classification Result Report lists "*Escherichia coli*" as best match with a log score value ≥ 2.00.	Passed
3. Blank value analysis of IVD Matrix HCCA-portioned by MALDI Biotyper	Classification Result Report lists "No Organism Identification Possible." with a log score value < 1.70.	Passed

Remarks:

1) IVD Matrix HCCA-portioned is shipped at ambient temperature. The expiry date on the package is valid for the product when stored in a refrigerator at 2°C to 8°C on arrival.

Bruker Daltonics GmbH & Co. KG
Fahrenheitstr. 4
28359 Bremen
Germany
Phone +49-421-2205-0

J. Vincenz
Quality Management

QC release date: 2026-05-05

Document Revision: 5017659D (DCR-13064)