# Certificate of Analysis

| | |
|---|---|
| Product Name : | **IVD Matrix HCCA-portioned** |
| **Part No.:** | **8290200** |
| **LOT No.:** | 6020226064 |
| **Use by[1]:** | 2028-05-05 |
| **Date of Production:** | 2026-05-06 |
| **Storage Conditions :** | **Store in a refrigerator at 2°C to 8°C on arrival.** |

| Test | Specification | LOT Results |
|---|---|---|
| **1. Amount of HCCA content** | AQL-value=1.0<br>Inspection level II<br>U=2.8 mg // L=2.2 mg<br>Sampling in compliance of DIN ISO 3951-1 | Passed |
| **2. MALDI Biotyper identification of a 1:20 diluted IVD BTS** | Classification Result Report lists "*Escherichia coli*" as best match with a log score value ≥ 2.00. | Passed |
| **3. Blank value analysis of IVD Matrix HCCA-portioned by MALDI Biotyper** | Classification Result Report lists "No Organism Identification Possible." with a log score value < 1.70. | Passed |

**Remarks:**

1) IVD Matrix HCCA-portioned is shipped at ambient temperature. The expiry date on the package is valid for the product when stored in a refrigerator at 2°C to 8°C on arrival.

Bruker Daltonics GmbH & Co. KG
Fahrenheitstr. 4
28359 Bremen
Germany
Phone +49-421-2205-0

J. Vincenz
Quality Management

QC release date: 2026-06-02

Document Revision: 5017659D (DCR-13064)