CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality Management System Regulation, ISO 13485, European IVD Directive, the Canadian Medical Devices Regulations (CMDR) and the China National Medical Products Administration's Quality System Requirements.

Product Name: Xpert® CT/NG

Catalogue Part No.: GXCT/NGX-CE-10

Kit Lot No.: 1001540526

Cartridge Lot No.: 39005

Kit Expiration Date: 2028-06-18

Legal Manufacturer
Cepheid AB
Röntgenvägen 5
SE-17154 Solna
Sweden

Manufacturing Facility

(●) Solna ( ) Sunnyvale

Cepheid AB
Röntgenvägen 5
SE-171 54 Solna
Sweden

Functional Testing

Test Description	*Acceptance Criteria*	*Test Result*
Positive	CT DETECTED; NG DETECTED	Passed
Negative	CT NOT DETECTED; NG NOT DETECTED	Passed

☑ If checked this document is produced electronically and valid without a wet signature.

Axel Eiji Ikegami Andersson
Axel Eiji Ikegami Andersson (07/07/2026 11:35:47 GMT+2)

2026-07-07

Signature of Quality Assurance **Date**

Name: Axel Andersson

Title: QA Analyst