Liofilchem®

Certificate of Analysis

Product	Batch	Production date	Expiration date
FLUCONAZOLE FLU 0.016-256 µg/mL	121525066	15.12.2025	2028.12.14

Ref.

92147 – 921470 – 921471

Antimicrobial Susceptibility Testing

Tested according to in-house methodology for antifungal agents

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
Candida krusei ATCC® 6258	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	128–≥256	256
Candida parapsilosis ATCC® 22019	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	1–8	2
Candida albicans ATCC® 90028	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.125–0.5	0.19

Batch Release

Approved

Date 18.12.2025 **Signature** Quality Control (This document has been established electronically and is valid without signature)

The results reported were obtained at the time of release.