Liofilchem®

Certificate of Analysis

Product	Batch	Production date	Expiration date
Vancomycin VA 0.016-256	082726070	27.08.2026	2029.08.26

Ref.

92057 – 920570 – 920571

Antimicrobial Susceptibility Testing

Tested according to current CLSI and EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC (mg/L)
Staphylococcus aureus ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.5–2 [a,b]	0.75
Enterococcus faecalis ATCC® 29212	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–4 [a,b]	2
Streptococcus pneumoniae ATCC® 49619	Mueller Hinton II Agar (Sheep blood 5%)	0.5 McFarland in broth	35 ± 2°C, 5% CO_2 20-24 h	0.12–0.5 [a]	0.38
Streptococcus pneumoniae ATCC® 49619	Mueller Hinton Fastidious Agar	0.5 McFarland in broth	35 ± 2°C, 5% CO_2 20-24 h	0.12–0.5 [b]	0.25

a)CLSI M100-Ed36
b)EUCAST QC tables v16.0

Batch Release

Approved

Date 31.08.2026 **Signature** Quality Control
(This document has been established electronically and is valid without signature)

The results reported were obtained at the time of release.