Liofilchem®

Certificate of Analysis

Product	Batch	Expiration date
Itraconazole ITC 8 µg	082626083	2029.08.25

Ref.

9139 – 9139/1 – 9139/2

Antimicrobial Susceptibility Testing

Tested according to in-house disc diffusion method for antifungal agents

Control strains	Medium	Inoculum	Incubation	Results Halo (mm)
Candida albicans ATCC® 90028	Mueller Hinton+ Glucose+ Methylen Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	36
Candida parapsilosis ATCC® 22019	Mueller Hinton+ Glucose+ Methylen Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	38
Candida tropicalis ATCC 750	Mueller Hinton+ Glucose+ Methylen Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	36
Candida krusei ATCC® 6258	Mueller Hinton+ Glucose+ Methylen Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	33

Batch Release

Approved

Date 04.09.2026 **Signature** Quality Control
(This document has been established electronically and is valid without signature)

The results reported were obtained at the time of release.