Liofilchem®

Certificate of Analysis

Product	Batch	Production date	Expiration date
Mecillinam MEC 0.016-256 µg/mL	070626095	06.07.2026	2029.07.05

Ref.

92017 – 920170 – 920171

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
Escherichia coli ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.03–0.25	0.094

Batch Release

Approved

Date 08.07.2026 **Signature** Quality Control
(This document has been established electronically and is valid without signature)

The results reported were obtained at the time of release.