Liofilchem®

Certificate of Analysis

Product	Batch	Production date	Expiration date
Cefazolin KZ 0.016-256 µg/mL	081826046	18.08.2026	2029.08.17

Ref.

92174 – 921740 – 921741

Antimicrobial Susceptibility Testing
Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
Staphylococcus aureus ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.25–1	0.38
Escherichia coli ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–4	1.5

Batch Release

Approved

Date 01.09.2026

Signature Quality Control
(This document has been established electronically and is valid without signature)

The results reported were obtained at the time of release.