Liofilchem®

Certificate of Analysis

Product	Batch	Expiration date
Bacitracin BA 10 IU	080626099	2029.08.05

Ref.

9051 – 9051/1 – 9051/2

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Results Zone (mm)
Escherichia coli ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	6
Staphylococcus aureus ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	20
Pseudomonas aeruginosa ATCC® 27853	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	6

Batch Release

Approved

Date 17.08.2026

Signature Quality Control (This document has been established electronically and is valid without signature)

The results reported were obtained at the time of release.