Certificate of Analysis

Product Name :	**IVD Bacterial Test Standard**
Part No.:	**8290190**
LOT No.:	6030326010
Use by[1)]:	2028-07-13
Date of Production:	2026-07-13
Storage Conditions :	**Store at -18 °C or below immediately on arrival.**

Test	Specification	LOT Results
Verification of the calibration points by MALDI-TOF MS	Within a mass range of ± 900 ppm	Passed
Relative intensity of reference protein No. 3 by MALDI-TOF MS	0.20 ± 0.05	Passed
Relative intensity of reference protein No. 4 by MALDI-TOF MS	0.22 ± 0.05	Passed
MALDI Biotyper identification	Classification Result Report lists "*Escherichia coli*" as best match.	Passed
	log(score) Value ≥ 2.2 of the best match	Passed

Remarks:

[1)] IVD BTS is shipped at ambient temperature but must be stored at –18°C or below immediately on arrival. The expiry date on the package is valid for the enclosed IVD BTS when stored at –18°C or below on arrival. Frozen, dissolved Bruker IVD BTS can be stored for up to 5 months at –18°C or below. See instruction for use at www.bruker.com/care.

Bruker Daltonics GmbH & Co. KG
Fahrenheitstr. 4
28359 Bremen
Germany
Phone +49-421-2205-0

Signiert von:
i.V. Sonja Lehmann
C2320FCF7496497...
Quality Management

QC release date: 2026-08-06

Document Revision: 5026845F (PKT-23051)